



**Government of West Bengal
Office of the Principal
Calcutta National Medical College
32, Gorachand Road, Kolkata-14**

Memo No. CNMC/ 9647 /P

Date: 28/02/2020

TENDER NOTICE

Principal, Calcutta National Medical College, Kolkata invites Sealed Tender from bonafied suppliers for the supply of **"Dental Chair with Compressor"** at the Department of Dentistry under Calcutta National Medical College, Kolkata. Each tender should contain two envelopes marked as "A" & "B" superscripting **"Tender for Dental Chair with Compressor" with Tender Notice No & Date & Name of the Bidder**. Envelope "A" should contain the documents as stated in Column 1 & Column 2 and technical bid. The Envelope "B" should contain price bid in duplicate.


The tender paper will have to be submitted on or before 06.03.2020 by 2.00 pm at receiving section. The sealed tender will be opened on 06.03.2020 at 2.30 pm at the chamber of the Principal, Calcutta National Medical College in presence of the intending tender bidders. The date of opening the price bid (Envelope B) will be notified later on. The Undersigned reserves the right to reject or accept any offer partly or wholly without assigning any reason thereof. The decision of the Undersigned in all respects is final and binding upon all concern.

General Terms & Conditions

1. Demand Draft in original showing deposit of Rs. 5000.00/- (Rupees five thousands only) as Earnest Money in favour of the undersigned. Without demand draft, the tender will be declared as invalid.
2. Attested Current Trade License, Income Tax Return submitted in last financial year, PAN Card, Professional Tax Registration certificate & GST Certificate (if applicable) must be attached with the tender, without such documents the tender will be declared as invalid.
3. Copy of evidence that the firm is acquainted/ equipped with the technical know-how of the equipment.
4. In case of failure to meet the technical specification of the equipment in the technical bid, the price bid will not be considered.
5. The bidder should duly sign all the pages of bidding documents with seal while submitting the technical part of the tender, i.e., Envelope A
6. Unsigned offers submitted by the bidder will be considered invalid.
7. Rates should be quoted both in figure & in words clearly (Indian Rupees) including all types of charges & taxes i.e., at the ultimate destination of the equipment. Price should be firm till execution of order.
8. The tenderers must quote basic rate along with 3 years comprehensive onsite warranty of entire system (including all spares & labour) from the date of completion of the satisfactory installation.
9. The lowest bidder must have to be given the physical demonstration as per specification after opening the tender as well as the price bid. Failure to provide demonstration within 3 days, the price bid will be treated cancelled and invalid. Under this circumstances, 2nd lowest bidder will be asked to provide physical demonstration in this regard
10. Approved tenderer is required to submit 10(ten) percent security deposit of the net value of the item at the time of delivery of the equipment.
11. In case of refusal/failure to supply, the EMD of the selected tenderer will be forfeited.
12. Packing & forwarding charges should be borne by the supplier. Rate should be inclusive of all.

13. Time of completion of delivery will be directed in order.
14. Service back up of the quoted equipment/instrument should be available in Kolkata.
15. Earnest Money of the unsuccessful bidders will be refunded in due course.
16. The Suppliers should have knowledge to supply the same kind of equipments to other Medical Colleges within the jurisdiction of Kolkata.
17. All suits arising out of the tender/ enquiry and subsequent purchase order/letter of acceptance (if any) are subjected to the jurisdiction in Kolkata only.

Enclosed:
Technical Specifications of Dental Chair with Compressor



28/02/2020

Principal

Calcutta National Medical College

Date: 28/02/2020

Memo No. CNMC/ 9647 /P


28/02/2020

Copy forwarded for information & necessary publication to:

1. The DDHS (E & S), Central Medical Stores,
2. The MSVP, CNMC&H
3. The Accounts Officer, CNMC
4. I.T. Cell, Swasthya Bhavan, for publication in the website
5. Notice Board, CNMC, Kolkata-14


28/02/2020

Principal

Calcutta National Medical College


28/02/2020

MEMO NO. C.N.M.C./STORE/9647/P. DT 28/02/2020

DENTAL CHAIR UNIT WITH COMPRESSOR - TECHNICAL SPECIFICATIONS

1. Should be electrically operated with Zero program.
2. LED operating light having 30000 LUX ON/OFF and intensity control by non-touch sensor with 3 directional movements $\pm 5\%$.
3. Should have auto water connection for spittoon and tumbler with auto filling sensor.
4. Should have high and low motorized vacuum suction with flow control valve
5. Should have an auto drain and auto flush system
6. Should have a multi functional foot control
7. Should have an Stainless Steel instrument tray and an X-ray viewer
8. Should have two air rotor points with hand piece and micro motor with hand piece and two sets 3 way syringe.
9. Should provide Dental operator's stool with back rest movement and hand rest.
10. Should be supplied with 0.75 HP oil free medical grade, noiseless compressor.
11. Should be operated in input supply 200 to 240 Volts, 50Hz.
12. Drainage system should be done by the company as turnkey work.
13. Company should ensure servicing for at least 3 consecutive years.

Krishna Padle Das

Indranil Senkar

Vishal Jain

Bhawati Chakraborty.

Adrite Roy Chowdhury.